

عنوان آزمایش : DHEA SO4-CLIA
عنوان دستگاه : abbot

سری ساخت : ۳۳۴۰۲۲۰۱

عنوان کنترل : abbot-1
شروع کنترل : ۱۴۰۰/۰۲/۰۱

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
35.59	33.999	32.408	30.817	29.226	27.635	26.044	1.591	5.163

تاریخ	مقدار
24/03/1400	28.1
25/03/1400	29.7
26/03/1400	28
27/03/1400	28.5
29/03/1400	29
30/03/1400	29.3
31/03/1400	30

تاریخ	مقدار
11/03/1400	28
12/03/1400	29.1
13/03/1400	29.8
17/03/1400	28.6
18/03/1400	29
19/03/1400	31.1
20/03/1400	28.3
22/03/1400	29
23/03/1400	31.1

تاریخ	مقدار
01/03/1400	29.6
02/03/1400	28.3
03/03/1400	29
04/03/1400	29.6
05/03/1400	28.4
06/03/1400	29
08/03/1400	30.9
09/03/1400	29.5
10/03/1400	29

Mean: 29.196	SD: 0.895	CV%: 3.066
--------------	-----------	------------

