

عنوان آزمایش : Free T4
عنوان دستگاه : simense

سری ساخت : ۴۰۳۸۲

عنوان کنترل : biorad II
شروع کنترل : ۱۳۹۹/۰۵/۲۳

| 3SD | 2SD | 1SD | MEAN | -1SD | -2SD | -3SD | SD | CV% |
|-------|-------|-------|-------|-------|------|-------|-------|-------|
| 3.135 | 3.042 | 2.949 | 2.856 | 2.763 | 2.67 | 2.577 | 0.093 | 3.241 |

| تاریخ | مقدار |
|------------|-------|
| 25/03/1400 | 2.88 |
| 26/03/1400 | 2.76 |
| 27/03/1400 | 2.82 |
| 29/03/1400 | 2.85 |
| 30/03/1400 | 2.85 |
| 31/03/1400 | 2.88 |

| تاریخ | مقدار |
|------------|-------|
| 11/03/1400 | 2.85 |
| 12/03/1400 | 3.03 |
| 13/03/1400 | 2.97 |
| 17/03/1400 | 2.97 |
| 18/03/1400 | 2.85 |
| 20/03/1400 | 3.05 |
| 22/03/1400 | 2.95 |
| 23/03/1400 | 2.95 |
| 24/03/1400 | 2.81 |

| تاریخ | مقدار |
|------------|-------|
| 01/03/1400 | 2.8 |
| 02/03/1400 | 2.75 |
| 03/03/1400 | 2.9 |
| 04/03/1400 | 2.75 |
| 05/03/1400 | 2.7 |
| 06/03/1400 | 2.84 |
| 08/03/1400 | 2.85 |
| 09/03/1400 | 2.81 |
| 10/03/1400 | 2.92 |

| | | |
|-------------|-----------|------------|
| Mean: 2.866 | SD: 0.089 | CV%: 3.098 |
|-------------|-----------|------------|

