

عنوان آزمایش : Insulin -CLIA
عنوان دستگاه : abbot

سری ساخت : ۳۳۴۰۲۲۰۱

عنوان کنترل : abbot-1
شروع کنترل : ۱۴۰۰/۰۲/۰۱

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
16.328	15.87	15.412	14.954	14.496	14.038	13.58	0.458	3.061

تاریخ	مقدار
24/03/1400	15.2
25/03/1400	15.1
26/03/1400	14.9
27/03/1400	14.7
29/03/1400	15.3
30/03/1400	15
31/03/1400	14.4

تاریخ	مقدار
11/03/1400	15.3
12/03/1400	15.6
13/03/1400	14.3
17/03/1400	15.5
18/03/1400	15.1
19/03/1400	15.3
20/03/1400	14.7
22/03/1400	15
23/03/1400	14.5

تاریخ	مقدار
01/03/1400	15.3
02/03/1400	14.9
03/03/1400	14.2
04/03/1400	14.6
05/03/1400	15.5
06/03/1400	14.7
08/03/1400	14.6
09/03/1400	15
10/03/1400	15.3

Mean: 14.96	SD: 0.392	CV%: 2.618
-------------	-----------	------------

