

عنوان آزمایش : Insulin -CLIA

عنوان دستگاه : siemens

سری ساخت : L1NC20134

عنوان کنترل :

شروع کنترل: 1400/04/05

II کنترل انسولین

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
51.999	49.666	47.333	45	42.667	40.334	38.001	2.333	7.3

تاریخ	مقدار
23/04/1400	41.2
24/04/1400	46.3
26/04/1400	48.3
28/04/1400	44.8
29/04/1400	46.1
31/04/1400	48.4

تاریخ	مقدار
14/04/1400	44.3
15/04/1400	42.7
16/04/1400	43.2
17/04/1400	44.7
19/04/1400	43.4
20/04/1400	42.2
21/04/1400	44.9
22/04/1400	45.3

تاریخ	مقدار
05/04/1400	44
06/04/1400	43
07/04/1400	45
08/04/1400	44.6
09/04/1400	43.2
10/04/1400	46.7
12/04/1400	47
13/04/1400	42.8

Mean: 44.641	SD: 1.912	CV%: 4.282
--------------	-----------	------------



7.3

SD 2.333

-3SD 38.001

-2SD 40.334


-1SD 42.667

Mean 45

1SD 47.333

2SD 49.666

3SD 51.999

چاپ 

siemens

