

عنوان آزمایش : PAPP-A-CLIA

عنوان دستگاه : simense

سری ساخت : 7942

عنوان کنترل :

randox I

شروع کنترل : 1400/07/18

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
1.755	1.682	1.609	1.536	1.463	1.39	1.317	0.073	4.723

تاریخ	مقدار
22/10/1400	1.6
23/10/1400	1.5
25/10/1400	1.5
28/10/1400	1.5
29/10/1400	1.6

تاریخ	مقدار
11/10/1400	1.6
12/10/1400	1.4
13/10/1400	1.5
14/10/1400	1.55
15/10/1400	1.6
18/10/1400	1.56
19/10/1400	1.52
20/10/1400	1.5

تاریخ	مقدار
01/10/1400	1.44
02/10/1400	1.59
04/10/1400	1.49
05/10/1400	1.62
06/10/1400	1.5
07/10/1400	1.54
08/10/1400	1.65
09/10/1400	1.52

Mean: 1.537	SD: 0.063	CV%: 4.069
-------------	-----------	------------



randox I

عنوان کنترل

اطلاعات

سری ساخت: 7942

آرشیو

شروع کنترل: 1400/07/18

مقدار

تاریخ

مقدار	تاریخ
1.44	1400/10/01
1.59	1400/10/02
1.49	1400/10/04
1.62	1400/10/05
1.5	1400/10/06
1.54	1400/10/07
1.65	1400/10/08
1.52	1400/10/09
1.6	1400/10/11
1.4	1400/10/12
1.5	1400/10/13
1.55	1400/10/14
1.6	1400/10/15
1.56	1400/10/18
1.52	1400/10/19
1.5	1400/10/20
1.6	1400/10/22
1.5	1400/10/23
1.5	1400/10/25
1.5	1400/10/28
1.6	1400/10/29

Mean: 1.537

SD: 0.063

CV%: 4.069

6σ

CV% 4.723

SD 0.073

-3SD 1.317

-2SD 1.39

-1SD 1.463

Mean 1.536

1SD 1.609

2SD 1.682

3SD 1.755

جاب

