

عنوان آزمایش : PAPP-A-CLIA

عنوان دستگاه : simense

MS-2 سری ساخت : 7462

عنوان کنترل :

randox III

شروع کنترل : 1399/10/22

| 3SD    | 2SD    | 1SD    | MEAN   | -1SD   | -2SD   | -3SD   | SD    | CV%   |
|--------|--------|--------|--------|--------|--------|--------|-------|-------|
| 17.368 | 16.853 | 16.338 | 15.823 | 15.308 | 14.793 | 14.278 | 0.515 | 3.255 |

| تاریخ      | مقدار |
|------------|-------|
| 24/04/1400 | 15.5  |
| 27/04/1400 | 16    |
| 28/04/1400 | 15.6  |
| 29/04/1400 | 16    |
| 31/04/1400 | 15.8  |

| تاریخ      | مقدار |
|------------|-------|
| 09/04/1400 | 16    |
| 10/04/1400 | 16.2  |
| 16/04/1400 | 16.2  |
| 17/04/1400 | 15.5  |
| 19/04/1400 | 16.5  |
| 21/04/1400 | 16.5  |
| 23/04/1400 | 15.6  |

| تاریخ      | مقدار |
|------------|-------|
| 01/04/1400 | 16.2  |
| 02/04/1400 | 15.2  |
| 03/04/1400 | 15.9  |
| 05/04/1400 | 15.5  |
| 06/04/1400 | 16    |
| 07/04/1400 | 16.2  |
| 08/04/1400 | 16.5  |

|              |           |            |
|--------------|-----------|------------|
| Mean: 15.942 | SD: 0.381 | CV%: 2.387 |
|--------------|-----------|------------|

