

randox I

عنوان کنترل

اطلاعات

سری ساخت: 7942

آرشیو

شروع کنترل: 1400/07/18

مقدار

تاریخ

1.54	1400/09/01
1.62	1400/09/02
1.65	1400/09/03
1.66	1400/09/04
1.45	1400/09/06
1.58	1400/09/07
1.48	1400/09/08
1.5	1400/09/09
1.69	1400/09/10
1.73	1400/09/11
1.69	1400/09/13
1.57	1400/09/14
1.76	1400/09/15
1.59	1400/09/16
1.48	1400/09/17
1.65	1400/09/18
1.57	1400/09/20
1.55	1400/09/21
1.71	1400/09/22
1.53	1400/09/23
1.62	1400/09/24
1.57	1400/09/25
1.53	1400/09/27
1.48	1400/09/28
1.43	1400/09/29

Mean: 1.578

SD: 0.097

CV%: 6.146

6σ

CV% 7.689 SD 0.118 -3SD 1.175 -2SD 1.293 -1SD 1.411 Mean 1.529 1SD 1.647 2SD 1.765 3SD 1.883

simense



عنوان آزمایش : PAPP-A-CLIA

عنوان دستگاه : simense

سری ساخت : 7942

عنوان کنترل :

شروع کنترل : 1400/07/18

randox I

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
1.883	1.765	1.647	1.529	1.411	1.293	1.175	0.118	7.689

تاریخ	مقدار
22/09/1400	1.71
23/09/1400	1.53
24/09/1400	1.62
25/09/1400	1.57
27/09/1400	1.53
28/09/1400	1.48
29/09/1400	1.43
30/09/1400	1.39

تاریخ	مقدار
11/09/1400	1.73
13/09/1400	1.69
14/09/1400	1.57
15/09/1400	1.76
16/09/1400	1.59
17/09/1400	1.48
18/09/1400	1.65
20/09/1400	1.57
21/09/1400	1.55

تاریخ	مقدار
01/09/1400	1.54
02/09/1400	1.62
03/09/1400	1.65
04/09/1400	1.66
06/09/1400	1.45
07/09/1400	1.58
08/09/1400	1.48
09/09/1400	1.5
10/09/1400	1.69

Mean: 1.578	SD: 0.097	CV%: 6.146
-------------	-----------	------------