

عنوان آزمایش : TSH
عنوان دستگاه : simense سری ساخت : ۴۰۳۸۲

عنوان کنترل : biorad II
شروع کنترل : ۱۳۹۹/۱۰/۱۵



3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
6.47	6.191	5.912	5.633	5.354	5.075	4.796	0.279	4.952

تاریخ	مقدار
22/03/1400	5.54
24/03/1400	5.63
25/03/1400	6.13
26/03/1400	5.24
27/03/1400	6.14
29/03/1400	5.68
30/03/1400	5.9

تاریخ	مقدار
10/03/1400	5.74
11/03/1400	5.23
12/03/1400	5.68
13/03/1400	5.77
17/03/1400	5.7
18/03/1400	5.66
19/03/1400	5.14
20/03/1400	5.79

تاریخ	مقدار
01/03/1400	5.65
02/03/1400	5.84
03/03/1400	5.6
04/03/1400	5.82
05/03/1400	5.96
06/03/1400	5.93
08/03/1400	6.03
09/03/1400	5.74

Mean: 5.719	SD: 0.259	CV%: 4.537
-------------	-----------	------------



عنوان آزمایش : TSH
عنوان دستگاه : simense سری ساخت : ۴۰۳۸۲

عنوان کنترل : biorad II
شروع کنترل : ۱۳۹۹/۱۰/۱۵

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
6.47	6.191	5.912	5.633	5.354	5.075	4.796	0.279	4.952

تاریخ	مقدار
22/03/1400	5.54
24/03/1400	5.63
25/03/1400	6.13
26/03/1400	5.24
27/03/1400	6.14
29/03/1400	5.68
30/03/1400	5.9

تاریخ	مقدار
10/03/1400	5.74
11/03/1400	5.23
12/03/1400	5.68
13/03/1400	5.77
17/03/1400	5.7
18/03/1400	5.66
19/03/1400	5.14
20/03/1400	5.79

تاریخ	مقدار
01/03/1400	5.65
02/03/1400	5.84
03/03/1400	5.6
04/03/1400	5.82
05/03/1400	5.96
06/03/1400	5.93
08/03/1400	6.03
09/03/1400	5.74

Mean: 5.719	SD: 0.259	CV%: 4.537
-------------	-----------	------------

