

عنوان آزمایش : Testosterone-CLIA
عنوان دستگاه : ABBOT

سری ساخت : ۳۳۴۰۲۲۰۱

عنوان کنترل : ABBOTT-1
شروع کنترل : ۱۴۰۰/۰۲/۰۱

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
23.5	22.025	20.55	19.075	17.6	16.125	14.65	1.475	7.733

تاریخ	مقدار
24/03/1400	19.3
25/03/1400	18.9
26/03/1400	18.7
27/03/1400	20.8
29/03/1400	22
30/03/1400	21.2
31/03/1400	20.4

تاریخ	مقدار
11/03/1400	19.8
12/03/1400	21.8
13/03/1400	20.2
17/03/1400	20.9
18/03/1400	21
19/03/1400	20.3
20/03/1400	22
22/03/1400	19
23/03/1400	19.6

تاریخ	مقدار
01/03/1400	19.9
02/03/1400	21
03/03/1400	20.7
04/03/1400	20.6
05/03/1400	19
06/03/1400	20.7
08/03/1400	20.3
09/03/1400	22
10/03/1400	18

Mean: 20.324	SD: 1.096	CV%: 5.392
--------------	-----------	------------

